

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100324\
 Data File : VY019789.D
 Acq On : 03 Oct 2024 15:41
 Operator : SY/MD
 Sample : P4270-03
 Misc : 4.58g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 IB-2-VOC

Quant Time: Oct 04 01:29:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	284160	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	553770	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	507270	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	196890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	175868	66.838	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.680%
35) Dibromofluoromethane	7.634	113	178831	56.957	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.920%
50) Toluene-d8	10.103	98	663524	57.135	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	114.260%
62) 4-Bromofluorobenzene	12.401	95	223213	49.987	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	99.980%
Target Compounds						
16) Acetone	3.873	43	34845	55.587	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100324\
Data File : VY019789.D
Acq On : 03 Oct 2024 15:41
Operator : SY/MD
Sample : P4270-03
Misc : 4.58g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
IB-2-VOC

Quant Time: Oct 04 01:29:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

