

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100421\
 Data File : VY006287.D
 Acq On : 04 Oct 2021 19:28
 Operator : SY/MD
 Sample : M4056-13
 Misc : 5.28G/5ML/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STOCK-PILE-PENINSULA-1B-11

Quant Time: Oct 05 06:38:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	109814	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	179400	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	163344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	67427	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	77480	65.92	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.84%
35) Dibromofluoromethane	7.722	113	53872	56.21	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.42%
50) Toluene-d8	10.185	98	222978	55.59	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.18%
62) 4-Bromofluorobenzene	12.483	95	65699	48.21	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.42%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100421\
Data File : VY006287.D
Acq On : 04 Oct 2021 19:28
Operator : SY/MD
Sample : M4056-13
Misc : 5.28G/5ML/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STOCK-PILE-PENINSULA-1B-11

Quant Time: Oct 05 06:38:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

