

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100421\
 Data File : VY006293.D
 Acq On : 04 Oct 2021 21:47
 Operator : SY/MD
 Sample : M4066-01
 Misc : 6.98G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NYPD-69TH-SOIL

Quant Time: Oct 05 06:39:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	117100	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	190032	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	173245	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	73920	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	81507	65.03	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.06%
35) Dibromofluoromethane	7.722	113	56735	55.88	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.76%
50) Toluene-d8	10.179	98	235764	55.49	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.98%
62) 4-Bromofluorobenzene	12.483	95	70608	48.92	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.84%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.954	59	3199	29.18	ug/l #	85
16) Acetone	3.954	43	6635	18.95	ug/l #	85
29) Tetrahydrofuran	7.350	42	92073	298.94	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100421\
Data File : VY006293.D
Acq On : 04 Oct 2021 21:47
Operator : SY/MD
Sample : M4066-01
Misc : 6.98G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NYPD-69TH-SOIL

Quant Time: Oct 05 06:39:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

