

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100421\
 Data File : VY006300.D
 Acq On : 05 Oct 2021 00:29
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 05 06:40:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	66	0.00
2 T	Dichlorodifluoromethane	0.507	0.453	10.7	60	0.00
3 P	Chloromethane	0.591	0.628	-6.3	69	0.00
4 C	Vinyl Chloride	0.655	0.709	-8.2#	72	0.00
5 T	Bromomethane	0.425	0.538	-26.6#	90	0.00
6 T	Chloroethane	0.385	0.472	-22.6	80	0.00
7 T	Trichlorofluoromethane	0.862	0.971	-12.6	75	0.00
8 T	Diethyl Ether	0.285	0.375	-31.6#	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.508	-10.4	74	0.00
10 T	Methyl Iodide	0.502	0.485	3.4	60	0.00
11 T	Tert butyl alcohol	0.047	0.066	-40.4#	92	0.00
12 CM	1,1-Dichloroethene	0.468	0.518	-10.7#	74	0.00
13 T	Acrolein	0.030	0.038	-26.7#	72	0.00
14 T	Allyl chloride	0.901	1.103	-22.4	80	0.00
15 T	Acrylonitrile	0.150	0.181	-20.7	77	0.00
16 T	Acetone	0.149	0.171	-14.8	64	0.00
17 T	Carbon Disulfide	1.566	1.745	-11.4	74	0.00
18 T	Methyl Acetate	0.504	0.783	-55.4#	99	0.00
19 T	Methyl tert-butyl Ether	1.368	1.638	-19.7	76	0.00
20 T	Methylene Chloride	0.582	0.694	-19.2	86	0.00
21 T	trans-1,2-Dichloroethene	0.519	0.549	-5.8	70	0.00
22 T	Diisopropyl ether	1.782	1.997	-12.1	73	0.00
23 T	Vinyl Acetate	1.068	1.190	-11.4	71	0.00
24 P	1,1-Dichloroethane	0.956	1.030	-7.7	71	0.00
25 T	2-Butanone	0.208	0.243	-16.8	69	0.00
26 T	2,2-Dichloropropane	0.875	0.836	4.5	63	0.00
27 T	cis-1,2-Dichloroethene	0.581	0.639	-10.0	72	0.00
28 T	Bromochloromethane	0.398	0.469	-17.8	79	0.00
29 T	Tetrahydrofuran	0.132	0.159	-20.5	76	0.00
30 C	Chloroform	0.954	1.082	-13.4#	75	0.00
31 T	Cyclohexane	1.014	0.911	10.2	62	0.00
32 T	1,1,1-Trichloroethane	0.881	0.971	-10.2	73	0.00
33 S	1,2-Dichloroethane-d4	0.535	0.610	-14.0	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	69	0.00
35 S	Dibromofluoromethane	0.267	0.270	-1.1	72	0.00
36 T	1,1-Dichloropropene	0.497	0.479	3.6	68	0.00
37 T	Ethyl Acetate	0.290	0.319	-10.0	74	0.00
38 T	Carbon Tetrachloride	0.525	0.539	-2.7	71	0.00
39 T	Methylcyclohexane	0.625	0.568	9.1	63	0.00
40 TM	Benzene	1.411	1.462	-3.6	71	0.00
41 T	Methacrylonitrile	0.188	0.210	-11.7	77	0.00
42 TM	1,2-Dichloroethane	0.455	0.530	-16.5	79	0.00
43 T	Isopropyl Acetate	0.556	0.630	-13.3	76	0.00
44 TM	Trichloroethene	0.365	0.382	-4.7	73	0.00
45 C	1,2-Dichloropropane	0.347	0.354	-2.0#	71	0.00
46 T	Dibromomethane	0.185	0.209	-13.0	76	0.00
47 T	Bromodichloromethane	0.476	0.529	-11.1	75	0.00
48 T	Methyl methacrylate	0.267	0.296	-10.9	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100421\
 Data File : VY006300.D
 Acq On : 05 Oct 2021 00:29
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 05 06:40:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	75	0.00
50 S	Toluene-d8	1.118	1.104	1.3	70	0.00
51 T	4-Methyl-2-Pentanone	0.289	0.344	-19.0	78	0.00
52 CM	Toluene	0.875	0.937	-7.1#	73	0.00
53 T	t-1,3-Dichloropropene	0.519	0.563	-8.5	74	0.00
54 T	cis-1,3-Dichloropropene	0.571	0.600	-5.1	71	0.00
55 T	1,1,2-Trichloroethane	0.260	0.284	-9.2	73	0.00
56 T	Ethyl methacrylate	0.396	0.461	-16.4	76	0.00
57 T	1,3-Dichloropropane	0.471	0.523	-11.0	75	0.00
58 T	2-Chloroethyl Vinyl ether	0.194	0.218	-12.4	76	0.00
59 T	2-Hexanone	0.206	0.246	-19.4	75	0.00
60 T	Dibromochloromethane	0.316	0.365	-15.5	77	0.00
61 T	1,2-Dibromoethane	0.246	0.272	-10.6	73	0.00
62 S	4-Bromofluorobenzene	0.380	0.390	-2.6	72	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	72	0.00
64 T	Tetrachloroethene	0.403	0.451	-11.9	82	0.00
65 PM	Chlorobenzene	0.991	1.016	-2.5	73	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.388	-6.3	76	0.00
67 C	Ethyl Benzene	1.867	1.903	-1.9#	73	0.00
68 T	m/p-Xylenes	0.709	0.732	-3.2	74	0.00
69 T	o-Xylene	0.669	0.710	-6.1	76	0.00
70 T	Styrene	1.137	1.218	-7.1	75	0.00
71 P	Bromoform	0.227	0.256	-12.8	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
73 T	Isopropylbenzene	3.582	3.606	-0.7	73	0.00
74 T	N-amyl acetate	1.112	1.192	-7.2	74	0.00
75 P	1,1,2,2-Tetrachloroethane	0.647	0.648	-0.2	72	0.00
76 T	1,2,3-Trichloropropane	0.486	0.486	0.0	75	0.00
77 T	Bromobenzene	0.840	0.883	-5.1	76	0.00
78 T	n-propylbenzene	4.409	4.383	0.6	73	0.00
79 T	2-Chlorotoluene	2.440	2.490	-2.0	75	0.00
80 T	1,3,5-Trimethylbenzene	2.995	3.072	-2.6	75	0.00
81 T	trans-1,4-Dichloro-2-butene	0.247	0.253	-2.4	72	0.00
82 T	4-Chlorotoluene	2.523	2.614	-3.6	76	0.00
83 T	tert-Butylbenzene	2.560	2.590	-1.2	74	0.00
84 T	1,2,4-Trimethylbenzene	2.957	3.078	-4.1	76	0.00
85 T	sec-Butylbenzene	3.869	3.857	0.3	72	0.00
86 T	p-Isopropyltoluene	3.265	3.323	-1.8	74	0.00
87 T	1,3-Dichlorobenzene	1.664	1.736	-4.3	77	0.00
88 T	1,4-Dichlorobenzene	1.640	1.707	-4.1	77	0.00
89 T	n-Butylbenzene	3.102	2.996	3.4	71	0.00
90 T	Hexachloroethane	0.593	0.600	-1.2	74	0.00
91 T	1,2-Dichlorobenzene	1.455	1.549	-6.5	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.123	0.140	-13.8	83	0.00
93 T	1,2,4-Trichlorobenzene	0.961	0.978	-1.8	74	0.00
94 T	Hexachlorobutadiene	0.627	0.608	3.0	71	0.00
95 T	Naphthalene	1.910	2.070	-8.4	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100421\
Data File : VY006300.D
Acq On : 05 Oct 2021 00:29
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 05 06:40:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.852	0.888	-4.2	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6