

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100423\
 Data File : VY015828.D
 Acq On : 04 Oct 2023 15:25
 Operator : SY/MD
 Sample : 04714-07
 Misc : 4.70g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: Oct 05 06:55:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	161722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	279813	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	258767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	119749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	108867	61.182	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.360%	
35) Dibromofluoromethane	7.716	113	89199	53.260	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.520%	
50) Toluene-d8	10.179	98	329051	57.946	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 115.900%	
62) 4-Bromofluorobenzene	12.477	95	112698	50.572	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.140%	
Target Compounds						
						Qvalue
16) Acetone	3.966	43	48070	50.480	ug/l	93
20) Methylene Chloride	4.710	84	43270	17.614	ug/l	91
25) 2-Butanone	6.990	43	22202	16.495	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100423\
Data File : VY015828.D
Acq On : 04 Oct 2023 15:25
Operator : SY/MD
Sample : 04714-07
Misc : 4.70g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: Oct 05 06:55:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

