

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100523\
 Data File : VY015849.D
 Acq On : 05 Oct 2023 14:56
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Oct 06 05:59:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 05:54:04 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	215264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	367998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	337277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	156171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	136674	64.534	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 129.060%		
35) Dibromofluoromethane	7.716	113	118453	56.471	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 112.940%		
50) Toluene-d8	10.179	98	434724	63.833	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 127.660%		
62) 4-Bromofluorobenzene	12.477	95	151016	54.249	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 108.500%		
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	34703	9.283	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100523\
Data File : VY015849.D
Acq On : 05 Oct 2023 14:56
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Oct 06 05:59:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 05:54:04 2023
Response via : Initial Calibration

