

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100623\
 Data File : VY015857.D
 Acq On : 06 Oct 2023 13:13
 Operator : SY/MD
 Sample : 04735-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT2662

Quant Time: Oct 07 07:25:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 05:54:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	35515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	53594	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	43300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	16963	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	18630	53.318	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.640%
35) Dibromofluoromethane	7.710	113	17374	56.873	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.740%
50) Toluene-d8	10.179	98	60904	61.405	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	122.820%
62) 4-Bromofluorobenzene	12.483	95	16999	41.930	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100623\
Data File : VY015857.D
Acq On : 06 Oct 2023 13:13
Operator : SY/MD
Sample : 04735-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT2662

Quant Time: Oct 07 07:25:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 05:54:04 2023
Response via : Initial Calibration

