

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100623\
 Data File : VY015859.D
 Acq On : 06 Oct 2023 14:00
 Operator : SY/MD
 Sample : 04736-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11991

Quant Time: Oct 07 07:26:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 05:54:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	189928	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	337615	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	312542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	144253	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	124437	66.594	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.180%
35) Dibromofluoromethane	7.716	113	109201	56.745	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.480%
50) Toluene-d8	10.179	98	397229	63.576	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	127.160%
62) 4-Bromofluorobenzene	12.483	95	137025	53.653	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100623\
Data File : VY015859.D
Acq On : 06 Oct 2023 14:00
Operator : SY/MD
Sample : 04736-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11991

Quant Time: Oct 07 07:26:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 05:54:04 2023
Response via : Initial Calibration

