

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100623\
 Data File : VY015862.D
 Acq On : 06 Oct 2023 15:11
 Operator : SY/MD
 Sample : 04765-03
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TT-2

Quant Time: Oct 07 07:27:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 05:54:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	221638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	392912	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	361053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	166422	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	146727	67.289	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.580%
35) Dibromofluoromethane	7.716	113	127097	56.750	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.500%
50) Toluene-d8	10.179	98	457597	62.931	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	125.860%
62) 4-Bromofluorobenzene	12.483	95	156945	52.804	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100623\
Data File : VY015862.D
Acq On : 06 Oct 2023 15:11
Operator : SY/MD
Sample : 04765-03
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TT-2

Quant Time: Oct 07 07:27:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 05:54:04 2023
Response via : Initial Calibration

