

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100821\
 Data File : VY006337.D
 Acq On : 07 Oct 2021 17:26
 Operator : SY/MD
 Sample : M4117-06
 Misc : 6.27G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-2(4.5-5.0)

Quant Time: Oct 08 07:13:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	56031	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	85451	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	74594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	33231	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	33794	56.35	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 112.70%	
35) Dibromofluoromethane	7.728	113	24071	52.73	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.46%	
50) Toluene-d8	10.185	98	105249	55.09	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.18%	
62) 4-Bromofluorobenzene	12.483	95	29235	45.04	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 90.08%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100821\
Data File : VY006337.D
Acq On : 07 Oct 2021 17:26
Operator : SY/MD
Sample : M4117-06
Misc : 6.27G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-2(4.5-5.0)

Quant Time: Oct 08 07:13:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

