

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100821\
 Data File : VY006345.D
 Acq On : 07 Oct 2021 20:31
 Operator : SY/MD
 Sample : M4117-18
 Misc : 4.36G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-5(6.5-7.0)

Quant Time: Oct 08 07:15:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	50557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	83069	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	80995	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	37985	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	37870	69.99	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.98%
35) Dibromofluoromethane	7.722	113	26863	60.53	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	121.06%
50) Toluene-d8	10.185	98	107657	57.96	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.92%
62) 4-Bromofluorobenzene	12.483	95	32851	52.06	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100821\
Data File : VY006345.D
Acq On : 07 Oct 2021 20:31
Operator : SY/MD
Sample : M4117-18
Misc : 4.36G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-5(6.5-7.0)

Quant Time: Oct 08 07:15:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

