



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100923\
Data File : VY015875.D
Acq On : 09 Oct 2023 10:54
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100523S.M
Title : SW846 8260
Last Update : Fri Oct 06 05:54:04 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.9	24885	PASS
75	95	30	60	52.3	68915	PASS
95	95	100	100	100.0	131723	PASS
96	95	5	9	6.5	8601	PASS
173	174	0.00	2	0.9	893	PASS
174	95	50	100	78.7	103720	PASS
175	174	5	9	7.6	7899	PASS
176	174	95	101	97.0	100576	PASS
177	176	5	9	6.9	6908	PASS