

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101021\
 Data File : VY006363.D
 Acq On : 08 Oct 2021 18:32
 Operator : SY/MD
 Sample : M4117-18RE
 Misc : 4.59G/5ML/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-5(6.5-7.0)RE

Quant Time: Oct 11 03:20:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	67645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	110836	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	104007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	47515	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	48882	67.517	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.040%
35) Dibromofluoromethane	7.722	113	34289	57.906	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.820%
50) Toluene-d8	10.185	98	138186	55.759	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.520%
62) 4-Bromofluorobenzene	12.483	95	41811	49.663	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.320%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.722	84	6542	6.098	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101021\
Data File : VY006363.D
Acq On : 08 Oct 2021 18:32
Operator : SY/MD
Sample : M4117-18RE
Misc : 4.59G/5ML/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-5(6.5-7.0)RE

Quant Time: Oct 11 03:20:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

