

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101022\
 Data File : VY010892.D
 Acq On : 10 Oct 2022 15:56
 Operator : KP/MD
 Sample : N5046-13
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 R-13

Quant Time: Oct 11 09:45:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	336825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	588615	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	511657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	221823	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	175632	51.442	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.880%
35) Dibromofluoromethane	7.716	113	177181	45.347	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.700%
50) Toluene-d8	10.179	98	667668	49.966	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.483	95	205616	38.543	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101022\
Data File : VY010892.D
Acq On : 10 Oct 2022 15:56
Operator : KP/MD
Sample : N5046-13
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
R-13

Quant Time: Oct 11 09:45:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

