

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101022\
 Data File : VY010894.D
 Acq On : 10 Oct 2022 16:41
 Operator : KP/MD
 Sample : N5046-06RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 R-6RE

Quant Time: Oct 11 09:45:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	8675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	14178	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	11785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	3418	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	5617	63.879	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.760%
35) Dibromofluoromethane	7.722	113	2679	28.465	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	56.940%
50) Toluene-d8	10.179	98	15431	47.943	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.880%
62) 4-Bromofluorobenzene	12.483	95	4001	30.374	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	60.740%
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	1132	7.087	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101022\
Data File : VY010894.D
Acq On : 10 Oct 2022 16:41
Operator : KP/MD
Sample : N5046-06RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
R-6RE

Quant Time: Oct 11 09:45:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

