

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101023\
 Data File : VY015896.D
 Acq On : 10 Oct 2023 14:54
 Operator : SY/MD
 Sample : 04800-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-100623

Quant Time: Oct 10 15:55:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 10 07:06:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	35136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	57054	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	49499	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	19665	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	19313	61.087	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.180%
35) Dibromofluoromethane	7.722	113	18581	58.388	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.780%
50) Toluene-d8	10.185	98	65799	62.737	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	125.480%
62) 4-Bromofluorobenzene	12.483	95	19200	45.071	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101023\
Data File : VY015896.D
Acq On : 10 Oct 2023 14:54
Operator : SY/MD
Sample : 04800-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-100623

Quant Time: Oct 10 15:55:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 10 07:06:18 2023
Response via : Initial Calibration

