

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101024\
 Data File : VY019852.D
 Acq On : 10 Oct 2024 14:30
 Operator : SY/MD
 Sample : P4364-03
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-4.25-4.83

Quant Time: Oct 11 01:42:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 10 05:30:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	13654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	23051	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	14978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	1653	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	7620	45.324	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.640%
35) Dibromofluoromethane	7.640	113	7362	48.399	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.800%
50) Toluene-d8	10.109	98	25117	44.714	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.420%
62) 4-Bromofluorobenzene	12.402	95	4382	21.590	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	43.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101024\
Data File : VY019852.D
Acq On : 10 Oct 2024 14:30
Operator : SY/MD
Sample : P4364-03
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-4.25-4.83

Quant Time: Oct 11 01:42:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 10 05:30:07 2024
Response via : Initial Calibration

