

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101024\
 Data File : VY019853.D
 Acq On : 10 Oct 2024 14:54
 Operator : SY/MD
 Sample : P4364-05
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-02-2.0-2.5

Quant Time: Oct 11 01:43:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 10 05:30:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	43403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	74515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	53089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	13172	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	23103	43.230	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.460%
35) Dibromofluoromethane	7.640	113	22748	46.262	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.520%
50) Toluene-d8	10.109	98	84117	46.324	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.640%
62) 4-Bromofluorobenzene	12.407	95	18596	28.342	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	56.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101024\
Data File : VY019853.D
Acq On : 10 Oct 2024 14:54
Operator : SY/MD
Sample : P4364-05
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-02-2.0-2.5

Quant Time: Oct 11 01:43:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 10 05:30:07 2024
Response via : Initial Calibration

