

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101024\
 Data File : VY019859.D
 Acq On : 10 Oct 2024 17:14
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Oct 11 01:45:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 10 05:30:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	288065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	580909	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	513386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	184804	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	178482	50.320	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.640%
35) Dibromofluoromethane	7.640	113	184763	48.199	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.400%
50) Toluene-d8	10.103	98	705657	49.848	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.700%
62) 4-Bromofluorobenzene	12.402	95	213410	41.722	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	83.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101024\
Data File : VY019859.D
Acq On : 10 Oct 2024 17:14
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Oct 11 01:45:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 10 05:30:07 2024
Response via : Initial Calibration

