

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101025\
 Data File : VY023440.D
 Acq On : 10 Oct 2025 15:37
 Operator : SY/MD
 Sample : Q3323-07
 Misc : 6.20g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SED-07-0-2

Quant Time: Oct 11 01:59:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	590270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1116913	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1047295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	372887	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	282258	57.184	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.360%
35) Dibromofluoromethane	7.640	113	366367	53.384	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.760%
50) Toluene-d8	10.109	98	1323044	51.766	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.540%
62) 4-Bromofluorobenzene	12.408	95	400451	47.458	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.920%
Target Compounds						
20) Methylene Chloride	4.616	84	38450	6.051	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101025\
Data File : VY023440.D
Acq On : 10 Oct 2025 15:37
Operator : SY/MD
Sample : Q3323-07
Misc : 6.20g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SED-07-0-2

Quant Time: Oct 11 01:59:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

