

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
 Data File : VY015914.D
 Acq On : 11 Oct 2023 10:48
 Operator : SY/MD
 Sample : VY1011SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1011SBL01

Quant Time: Oct 12 01:54:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 10 07:06:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	154865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	272397	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	248132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	112752	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	91149	65.411	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.820%
35) Dibromofluoromethane	7.722	113	87644	57.684	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.360%
50) Toluene-d8	10.179	98	318291	63.564	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	127.120%
62) 4-Bromofluorobenzene	12.483	95	108137	53.168	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
Data File : VY015914.D
Acq On : 11 Oct 2023 10:48
Operator : SY/MD
Sample : VY1011SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBL01

Quant Time: Oct 12 01:54:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 10 07:06:18 2023
Response via : Initial Calibration

