

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
 Data File : VY015918.D
 Acq On : 11 Oct 2023 13:00
 Operator : SY/MD
 Sample : 04817-04
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-241

Quant Time: Oct 12 01:56:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 10 07:06:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	153402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	269032	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	258338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	121452	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	90642	65.668	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	131.340%	
35) Dibromofluoromethane	7.716	113	86750	57.810	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	115.620%	
50) Toluene-d8	10.179	98	324455	65.606	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	131.220%	
62) 4-Bromofluorobenzene	12.483	95	116087	57.791	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	115.580%	
Target Compounds						
16) Acetone	3.954	43	4268	12.213	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
Data File : VY015918.D
Acq On : 11 Oct 2023 13:00
Operator : SY/MD
Sample : 04817-04
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-241

Quant Time: Oct 12 01:56:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 10 07:06:18 2023
Response via : Initial Calibration

