

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
 Data File : VY015926.D
 Acq On : 11 Oct 2023 16:08
 Operator : SY/MD
 Sample : 04839-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-101023

Quant Time: Oct 12 01:57:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 10 07:06:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	146222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	255461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	239439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	109107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	90471	68.762	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 137.520%		
35) Dibromofluoromethane	7.716	113	84557	59.342	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 118.680%		
50) Toluene-d8	10.179	98	304379	64.816	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 129.640%		
62) 4-Bromofluorobenzene	12.483	95	103251	54.131	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 108.260%		
Target Compounds						
20) Methylene Chloride	4.716	84	22838	8.944	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
Data File : VY015926.D
Acq On : 11 Oct 2023 16:08
Operator : SY/MD
Sample : 04839-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-02-101023

Quant Time: Oct 12 01:57:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 10 07:06:18 2023
Response via : Initial Calibration

