

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
 Data File : VY015916.D
 Acq On : 11 Oct 2023 11:43
 Operator : SY/MD
 Sample : VY1011SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1011SBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/12/2023
 Supervised By : Mahesh Dadoda 10/12/2023

Quant Time: Oct 12 01:55:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 10 07:06:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	178699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	293658	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	262054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	130358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	80859	50.288	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 100.580%			
35) Dibromofluoromethane	7.722	113	84104	51.347	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 102.700%			
50) Toluene-d8	10.185	98	284322	52.669	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 105.340%			
62) 4-Bromofluorobenzene	12.483	95	109098	49.757	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 99.520%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	14311	20.022	ug/l	93
3) Chloromethane	2.113	50	12533	17.143	ug/l	98
4) Vinyl Chloride	2.253	62	16937	18.164	ug/l	99
5) Bromomethane	2.644	94	14055	18.382	ug/l	97
6) Chloroethane	2.790	64	12869	17.576	ug/l	97
7) Trichlorofluoromethane	3.125	101	39085	19.754	ug/l	96
8) Diethyl Ether	3.534	74	15750	18.816	ug/l	89
9) 1,1,2-Trichlorotrifluo...	3.906	101	27085	20.171	ug/l	94
10) Methyl Iodide	4.095	142	18833	15.592	ug/l	96
11) Tert butyl alcohol	4.960	59	20887	57.651	ug/l #	81
12) 1,1-Dichloroethene	3.875	96	19509	18.889	ug/l	89
13) Acrolein	3.735	56	8738	88.913	ug/l	97
14) Allyl chloride	4.485	41	35579	18.683	ug/l	90
15) Acrylonitrile	5.168	53	53070	92.200	ug/l	98
16) Acetone	3.948	43	61290	83.266	ug/l	94
17) Carbon Disulfide	4.198	76	19155	14.457	ug/l	95
18) Methyl Acetate	4.479	43	36431	16.713	ug/l #	90
19) Methyl tert-butyl Ether	5.228	73	90014	19.232	ug/l	99
20) Methylene Chloride	4.722	84	37747	16.202	ug/l	89
21) trans-1,2-Dichloroethene	5.228	96	22245	18.561	ug/l	94
22) Diisopropyl ether	6.131	45	101978	20.021	ug/l #	88
23) Vinyl Acetate	6.064	43	267755	95.598	ug/l #	94
24) 1,1-Dichloroethane	6.027	63	55424	19.754	ug/l	95
25) 2-Butanone	6.990	43	80215	90.942	ug/l	95
26) 2,2-Dichloropropane	6.990	77	50140	19.216	ug/l	97
27) cis-1,2-Dichloroethene	6.990	96	34912	19.737	ug/l	92
28) Bromochloromethane	7.338	49	33813	20.910	ug/l	92
29) Tetrahydrofuran	7.356	42	42617	88.320	ug/l	91
30) Chloroform	7.509	83	63867	20.408	ug/l	99
31) Cyclohexane	7.789	56	30275	17.580	ug/l #	79
32) 1,1,1-Trichloroethane	7.704	97	51550	20.170	ug/l	98
36) 1,1-Dichloropropene	7.923	75	33742	18.600	ug/l	99
37) Ethyl Acetate	7.082	43	29159	18.716	ug/l	96
38) Carbon Tetrachloride	7.905	117	43727	20.071	ug/l	96
39) Methylcyclohexane	9.185	83	33305	18.074	ug/l	96
40) Benzene	8.167	78	110759	19.428	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
 Data File : VY015916.D
 Acq On : 11 Oct 2023 11:43
 Operator : SY/MD
 Sample : VY1011SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1011SBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/12/2023
 Supervised By : Mahesh Dadoda 10/12/2023

Quant Time: Oct 12 01:55:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 10 07:06:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.313	41	18288m	18.872	ug/l	
42) 1,2-Dichloroethane	8.240	62	37690	19.394	ug/l	92
43) Isopropyl Acetate	8.277	43	54259	18.814	ug/l	95
44) Trichloroethene	8.941	130	33079	20.052	ug/l	93
45) 1,2-Dichloropropane	9.216	63	34238	19.967	ug/l	100
46) Dibromomethane	9.307	93	19947	19.547	ug/l	96
47) Bromodichloromethane	9.502	83	51753	20.392	ug/l	95
48) Methyl methacrylate	9.295	41	23974	18.824	ug/l	88
49) 1,4-Dioxane	9.301	88	6468	364.626	ug/l #	45
51) 4-Methyl-2-Pentanone	10.075	43	162931	94.429	ug/l	92
52) Toluene	10.246	92	75685	20.000	ug/l	95
53) t-1,3-Dichloropropene	10.465	75	50483	19.676	ug/l	98
54) cis-1,3-Dichloropropene	9.929	75	54469	19.544	ug/l #	90
55) 1,1,2-Trichloroethane	10.648	97	33374	20.540	ug/l	98
56) Ethyl methacrylate	10.514	69	42740	19.375	ug/l #	84
57) 1,3-Dichloropropane	10.795	76	52176	19.917	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.783	63	130175	96.818	ug/l	94
59) 2-Hexanone	10.837	43	122086	93.526	ug/l	89
60) Dibromochloromethane	10.990	129	39133	20.090	ug/l	99
61) 1,2-Dibromoethane	11.093	107	28911	19.482	ug/l	99
64) Tetrachloroethene	10.721	164	34125	19.954	ug/l	96
65) Chlorobenzene	11.520	112	93047	20.356	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.593	131	38699	20.714	ug/l	98
67) Ethyl Benzene	11.593	91	158782	20.248	ug/l	100
68) m/p-Xylenes	11.703	106	119336	40.386	ug/l	96
69) o-Xylene	12.032	106	59699	20.192	ug/l	98
70) Styrene	12.044	104	106377	20.622	ug/l	96
71) Bromoform	12.209	173	26048	20.021	ug/l #	98
73) Isopropylbenzene	12.331	105	167050	20.315	ug/l	99
74) N-amyl acetate	12.148	43	49888	19.251	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.581	83	42309	19.843	ug/l	98
76) 1,2,3-Trichloropropane	12.636	75	28421m	18.676	ug/l	
77) Bromobenzene	12.611	156	40613	20.142	ug/l	98
78) n-propylbenzene	12.672	91	204206	20.650	ug/l	100
79) 2-Chlorotoluene	12.758	91	116213	20.631	ug/l	99
80) 1,3,5-Trimethylbenzene	12.812	105	140125	20.675	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.380	75	12428	18.110	ug/l	96
82) 4-Chlorotoluene	12.855	91	120471	20.523	ug/l	99
83) tert-Butylbenzene	13.075	119	132534	20.954	ug/l	97
84) 1,2,4-Trimethylbenzene	13.123	105	140296	20.556	ug/l	98
85) sec-Butylbenzene	13.258	105	192389	20.740	ug/l	100
86) p-Isopropyltoluene	13.373	119	160172	20.873	ug/l	98
87) 1,3-Dichlorobenzene	13.367	146	83468	20.561	ug/l	100
88) 1,4-Dichlorobenzene	13.447	146	83539	20.435	ug/l	99
89) n-Butylbenzene	13.696	91	154101	21.221	ug/l	97
90) Hexachloroethane	13.965	117	31547	20.485	ug/l	92
91) 1,2-Dichlorobenzene	13.739	146	78085	20.694	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.361	75	6905	18.440	ug/l	83
93) 1,2,4-Trichlorobenzene	15.007	180	49275	20.376	ug/l	98
94) Hexachlorobutadiene	15.117	225	26882	20.759	ug/l	99
95) Naphthalene	15.239	128	108337	19.365	ug/l	99
96) 1,2,3-Trichlorobenzene	15.428	180	43917	20.415	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
Data File : VY015916.D
Acq On : 11 Oct 2023 11:43
Operator : SY/MD
Sample : VY1011SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01


Manual Integrations
APPROVED

Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023

Quant Time: Oct 12 01:55:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 10 07:06:18 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:168 Resp: 178699
Ion Ratio Lower Upper
168 100
99 53.4 43.4 65.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#2
Dichlorodifluoromethane
Concen: 20.022 ug/l
RT: 1.906 min Scan# 14
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 85 Resp: 14311
Ion Ratio Lower Upper
85 100
87 35.2 15.6 46.8









#3
Chloromethane
Concen: 17.143 ug/l
RT: 2.113 min Scan# 49
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 50 Resp: 1253
Ion Ratio Lower Upper
50 100
52 32.7 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#4
Vinyl Chloride
Concen: 18.164 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 62 Resp: 16937
Ion Ratio Lower Upper
62 100
64 30.1 24.3 36.5





#5
Bromomethane
Concen: 18.382 ug/l
RT: 2.644 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 94 Resp: 1405
Ion Ratio Lower Upper
94 100
96 98.0 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#6
Chloroethane
Concen: 17.576 ug/l
RT: 2.790 min Scan# 159
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 64 Resp: 12869
Ion Ratio Lower Upper
64 100
66 28.9 24.4 36.6





#7
 Trichlorofluoromethane
 Concen: 19.754 ug/l
 RT: 3.125 min Scan# 214
 Delta R.T. 0.000 min
 Lab File: VY015916.D
 Acq: 11 Oct 2023 11:43




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1011SBSD01




Tgt Ion:101 Resp: 3908
 Ion Ratio Lower Upper
 101 100
 103 61.1 51.4 77.0

Manual Integrations
 APPROVED



Reviewed By :John Carlone 10/12/2023
 Supervised By :Mahesh Dadoda 10/12/2023




#8
 Diethyl Ether
 Concen: 18.816 ug/l
 RT: 3.534 min Scan# 281
 Delta R.T. 0.000 min
 Lab File: VY015916.D
 Acq: 11 Oct 2023 11:43



Tgt Ion: 74 Resp: 15750
 Ion Ratio Lower Upper
 74 100
 45 91.7 51.5 154.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.171 ug/l
RT: 3.906 min Scan# 342
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:101 Resp: 27085
Ion Ratio Lower Upper
101 100
85 43.7 36.6 55.0
151 76.8 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#10
Methyl Iodide
Concen: 15.592 ug/l
RT: 4.095 min Scan# 373
Delta R.T. 0.001 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:142 Resp: 18833
Ion Ratio Lower Upper
142 100
127 43.4 37.6 56.4
141 14.3 11.8 17.8









#11
Tert butyl alcohol
Concen: 57.651 ug/l
RT: 4.960 min Scan# 516
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 59 Resp: 2088
Ion Ratio Lower Upper
59 100
57 9.9 3.0 4.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#12
1,1-Dichloroethene
Concen: 18.889 ug/l
RT: 3.875 min Scan# 337
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 96 Resp: 19509
Ion Ratio Lower Upper
96 100
61 152.7 136.4 204.6
98 65.0 49.4 74.2





#13
Acrolein
Concen: 88.913 ug/l
RT: 3.735 min Scan# 314
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 56 Resp: 8738
Ion Ratio Lower Upper
56 100
55 68.7 57.0 85.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#14
Allyl chloride
Concen: 18.683 ug/l
RT: 4.485 min Scan# 437
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 41 Resp: 35579
Ion Ratio Lower Upper
41 100
39 64.3 58.0 87.0
76 39.9 27.4 41.2





#15
Acrylonitrile
Concen: 92.200 ug/l
RT: 5.168 min Scan# 549
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 53 Resp: 53070
Ion Ratio Lower Upper
53 100
52 81.8 66.9 100.3
51 36.3 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#16
Acetone
Concen: 83.266 ug/l
RT: 3.948 min Scan# 349
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 43 Resp: 61290
Ion Ratio Lower Upper
43 100
58 31.7 22.7 34.1





#17
Carbon Disulfide
Concen: 14.457 ug/l
RT: 4.198 min Scan# 390
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01




Tgt Ion: 76 Resp: 1915
Ion Ratio Lower Upper
76 100
78 10.4 7.0 10.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023




#18
Methyl Acetate
Concen: 16.713 ug/l
RT: 4.479 min Scan# 436
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43



Tgt Ion: 43 Resp: 36431
Ion Ratio Lower Upper
43 100
74 26.6 17.4 26.0#









#19
Methyl tert-butyl Ether
Concen: 19.232 ug/l
RT: 5.228 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 73 Resp: 90014
Ion Ratio Lower Upper
73 100
57 21.2 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#20
Methylene Chloride
Concen: 16.202 ug/l
RT: 4.722 min Scan# 476
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 84 Resp: 37747
Ion Ratio Lower Upper
84 100
49 114.1 106.9 160.3
51 36.6 33.2 49.8
86 65.3 51.0 76.6





#21
trans-1,2-Dichloroethene
Concen: 18.561 ug/l
RT: 5.228 min Scan# 511
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01




Tgt Ion: 96 Resp: 22249
Ion Ratio Lower Upper
96 100
61 127.5 110.2 165.4
98 64.0 51.8 77.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023




#22
Diisopropyl ether
Concen: 20.021 ug/l
RT: 6.131 min Scan# 707
Delta R.T. 0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43



Tgt Ion: 45 Resp: 101978
Ion Ratio Lower Upper
45 100
43 51.8 52.2 78.4#
87 27.2 21.0 31.6
59 12.1 8.9 13.3









#23
 Vinyl Acetate
 Concen: 95.598 ug/l
 RT: 6.064 min Scan# 696
 Delta R.T. 0.000 min
 Lab File: VY015916.D
 Acq: 11 Oct 2023 11:43

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1011SBSD01

Tgt Ion: 43 Resp: 26775
 Ion Ratio Lower Upper
 43 100
 86 11.5 7.5 11.3




Manual Integrations
 APPROVED


Reviewed By :John Carlone 10/12/2023
 Supervised By :Mahesh Dadoda 10/12/2023



#24
 1,1-Dichloroethane
 Concen: 19.754 ug/l
 RT: 6.027 min Scan# 690
 Delta R.T. 0.000 min
 Lab File: VY015916.D
 Acq: 11 Oct 2023 11:43

Tgt Ion: 63 Resp: 55424
 Ion Ratio Lower Upper
 63 100
 98 8.0 3.1 9.4
 100 5.2 1.8 5.5





#25

2-Butanone

Concen: 90.942 ug/l

RT: 6.990 min Scan# 848

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01

Tgt Ion: 43 Resp: 80215

Ion Ratio Lower Upper

43 100





72 28.0 20.5 30.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#26

2,2-Dichloropropane

Concen: 19.216 ug/l

RT: 6.990 min Scan# 848

Delta R.T. 0.000 min

Lab File: VY015916.D




Acq: 11 Oct 2023 11:43


Tgt Ion: 77 Resp: 50140

Ion Ratio Lower Upper

77 100


97 23.3 10.9 32.9





#27
 cis-1,2-Dichloroethene
 Concen: 19.737 ug/l
 RT: 6.990 min Scan# 848
 Delta R.T. 0.000 min
 Lab File: VY015916.D
 Acq: 11 Oct 2023 11:43




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1011SBSD01




Tgt Ion: 96 Resp: 34912
 Ion Ratio Lower Upper
 96 100
 61 137.6 0.0 301.4
 98 64.3 0.0 130.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 10/12/2023
 Supervised By :Mahesh Dadoda 10/12/2023




#28
 Bromochloromethane
 Concen: 20.910 ug/l
 RT: 7.338 min Scan# 905
 Delta R.T. 0.000 min
 Lab File: VY015916.D
 Acq: 11 Oct 2023 11:43



Tgt Ion: 49 Resp: 33813
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 4.6
 130 82.9 60.7 91.1





#29

Tetrahydrofuran

Concen: 88.320 ug/l

RT: 7.356 min Scan# 908

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01





Tgt Ion:	42	Resp:	4261
Ion Ratio	Lower	Upper	
42	100		
72	47.0	32.8	49.2
71	43.8	31.0	46.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#30

Chloroform

Concen: 20.408 ug/l




RT: 7.509 min Scan# 933


Delta R.T. -0.006 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43


Tgt Ion:	83	Resp:	63867
Ion Ratio	Lower	Upper	
83	100		
85	63.1	49.7	74.5





#31
Cyclohexane
Concen: 17.580 ug/l
RT: 7.789 min Scan# 911
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 56 Resp: 3027
Ion Ratio Lower Upper
56 100
69 48.3 27.8 41.6
84 101.5 66.4 99.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#32
1,1,1-Trichloroethane
Concen: 20.170 ug/l
RT: 7.704 min Scan# 965
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 97 Resp: 51550
Ion Ratio Lower Upper
97 100
99 63.7 50.8 76.2
61 41.2 35.1 52.7





#33

1,2-Dichloroethane-d4

Concen: 50.288 ug/l

RT: 8.149 min Scan# 1038

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01

Tgt Ion: 65 Resp: 80859

Ion Ratio Lower Upper

65 100





67 53.7 0.0 101.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.006 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43




Tgt Ion:114 Resp: 293658


Ion Ratio Lower Upper

114 100

63 18.9 0.0 42.4

88 15.1 0.0 30.6









#35
Dibromofluoromethane
Concen: 51.347 ug/l
RT: 7.722 min Scan# 968
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:113 Resp: 84104
Ion Ratio Lower Upper
113 100
111 103.9 81.3 121.9
192 19.5 16.9 25.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#36
1,1-Dichloropropene
Concen: 18.600 ug/l
RT: 7.923 min Scan# 1001
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 75 Resp: 33742
Ion Ratio Lower Upper
75 100
110 35.6 17.8 53.5
77 31.8 24.4 36.6





#37
Ethyl Acetate
Concen: 18.716 ug/l
RT: 7.082 min Scan# 863
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 43 Resp: 2915
Ion Ratio Lower Upper
43 100
61 13.8 10.1 15.1
70 10.4 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#38
Carbon Tetrachloride
Concen: 20.071 ug/l
RT: 7.905 min Scan# 998
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion:117 Resp: 43727
Ion Ratio Lower Upper
117 100
119 98.5 76.4 114.6
121 29.2 25.8 38.6





#39
Methylcyclohexane
Concen: 18.074 ug/l
RT: 9.185 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 83 Resp: 3330
Ion Ratio Lower Upper
83 100
55 73.2 63.0 94.4
98 46.6 37.4 56.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#40
Benzene
Concen: 19.428 ug/l
RT: 8.167 min Scan# 1041
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 78 Resp: 110759
Ion Ratio Lower Upper
78 100
77 23.8 19.0 28.6





#41
Methacrylonitrile
Concen: 18.872 ug/l m
RT: 7.313 min Scan# 904
Delta R.T. -0.018 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 41 Resp: 1828
Ion Ratio Lower Upper
41 100
39 50.6 52.1 78.1
67 74.3 66.2 99.2
52 27.8 29.9 44.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#42
1,2-Dichloroethane
Concen: 19.394 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 62 Resp: 37690
Ion Ratio Lower Upper
62 100
98 10.8 0.0 15.8





#43

Isopropyl Acetate

Concen: 18.814 ug/l

RT: 8.277 min Scan# 1059

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01





Tgt Ion:	43	Resp:	5425
Ion Ratio	Lower	Upper	
43	100		
61	28.6	21.1	31.7
87	14.5	9.8	14.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#44

Trichloroethene

Concen: 20.052 ug/l




RT: 8.941 min Scan# 1168


Delta R.T. -0.006 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Tgt Ion:	130	Resp:	33079
Ion Ratio	Lower	Upper	
130	100		
95	91.1	0.0	196.0





#45

1,2-Dichloropropane

Concen: 19.967 ug/l

RT: 9.216 min Scan# 1213

Delta R.T. -0.006 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01

Tgt Ion: 63 Resp: 34238

Ion Ratio Lower Upper

63 100





65 29.8 23.8 35.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#46

Dibromomethane

Concen: 19.547 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43




Tgt Ion: 93 Resp: 19947


Ion Ratio Lower Upper

93 100

95 84.0 66.6 100.0

174 99.3 84.8 127.2









#47
Bromodichloromethane
Concen: 20.392 ug/l
RT: 9.502 min Scan# 1260
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 83 Resp: 5175
Ion Ratio Lower Upper
83 100
85 66.7 50.1 75.1
127 10.4 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#48
Methyl methacrylate
Concen: 18.824 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 41 Resp: 23974
Ion Ratio Lower Upper
41 100
69 91.7 64.5 96.7
39 54.5 50.8 76.2





#49
1,4-Dioxane
Concen: 364.626 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 88 Resp: 6468
Ion Ratio Lower Upper
88 100
43 27.1 0.0 0.0
58 72.0 30.9 46.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#50
Toluene-d8
Concen: 52.669 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 98 Resp: 284322
Ion Ratio Lower Upper
98 100
100 65.5 50.1 75.1









#51
4-Methyl-2-Pentanone
Concen: 94.429 ug/l
RT: 10.075 min Scan# 1354
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 43 Resp: 16293
Ion Ratio Lower Upper
43 100
58 41.3 29.0 43.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#52
Toluene
Concen: 20.000 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 92 Resp: 75685
Ion Ratio Lower Upper
92 100
91 172.6 143.8 215.6









#53
t-1,3-Dichloropropene
Concen: 19.676 ug/l
RT: 10.465 min Scan# 1419
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 75 Resp: 50483
Ion Ratio Lower Upper
75 100
77 30.3 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#54
cis-1,3-Dichloropropene
Concen: 19.544 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 75 Resp: 54469
Ion Ratio Lower Upper
75 100
77 30.1 24.2 36.4
39 41.8 42.2 63.4#





#55
1,1,2-Trichloroethane
Concen: 20.540 ug/l
RT: 10.648 min Scan# 1448
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01




Tgt Ion:	97	Resp:	33374
Ion	Ratio	Lower	Upper
97	100		
83	87.4	72.0	108.0
85	57.4	45.4	68.0
99	61.0	50.3	75.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023




#56
Ethyl methacrylate
Concen: 19.375 ug/l
RT: 10.514 min Scan# 1426
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Tgt Ion:	69	Resp:	42740
Ion	Ratio	Lower	Upper
69	100		
41	59.5	57.6	86.4
39	30.9	34.1	51.1





#57
1,3-Dichloropropane
Concen: 19.917 ug/l
RT: 10.795 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 76 Resp: 52170
Ion Ratio Lower Upper
76 100
78 32.0 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#58
2-Chloroethyl Vinyl ether
Concen: 96.818 ug/l
RT: 9.783 min Scan# 1306
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 63 Resp: 130175
Ion Ratio Lower Upper
63 100
106 29.3 21.2 31.8









#59
2-Hexanone
Concen: 93.526 ug/l
RT: 10.837 min Scan# 1479
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 43 Resp: 122086
Ion Ratio Lower Upper
43 100
58 57.6 25.1 75.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#60
Dibromochloromethane
Concen: 20.090 ug/l
RT: 10.990 min Scan# 1504
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:129 Resp: 39133
Ion Ratio Lower Upper
129 100
127 76.3 38.5 115.5









#61
1,2-Dibromoethane
Concen: 19.482 ug/l
RT: 11.093 min Scan# 1521
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:107 Resp: 2891
Ion Ratio Lower Upper
107 100
109 93.2 75.6 113.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#62
4-Bromofluorobenzene
Concen: 49.757 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 95 Resp: 109098
Ion Ratio Lower Upper
95 100
174 83.8 0.0 178.2
176 81.7 0.0 173.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion:117 Resp: 262054
Ion Ratio Lower Upper
117 100
82 52.6 43.4 65.0
119 31.3 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#64
Tetrachloroethene
Concen: 19.954 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:164 Resp: 34125
Ion Ratio Lower Upper
164 100
166 121.8 103.0 154.6
129 92.9 77.2 115.8
131 85.0 69.6 104.4









#65
Chlorobenzene
Concen: 20.356 ug/l
RT: 11.520 min Scan# 11591
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:112 Resp: 9304
Ion Ratio Lower Upper
112 100
114 32.3 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#66
1,1,1,2-Tetrachloroethane
Concen: 20.714 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:131 Resp: 38699
Ion Ratio Lower Upper
131 100
133 95.0 47.5 142.6
119 66.2 31.6 94.7









#67
Ethyl Benzene
Concen: 20.248 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 91 Resp: 158782
Ion Ratio Lower Upper
91 100
106 30.6 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#68
m/p-Xylenes
Concen: 40.386 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:106 Resp: 119336
Ion Ratio Lower Upper
106 100
91 200.8 166.1 249.1









#69
o-Xylene
Concen: 20.192 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:106 Resp: 5969
Ion Ratio Lower Upper
106 100
91 212.2 108.1 324.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#70
Styrene
Concen: 20.622 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:104 Resp: 106377
Ion Ratio Lower Upper
104 100
78 47.5 41.5 62.3
103 54.4 44.6 67.0





#71

Bromoform

Concen: 20.021 ug/l

RT: 12.209 min Scan# 1704

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01





Tgt Ion:173	Resp:	2604
Ion Ratio	Lower	Upper
173	100	
175	47.6	24.6 73.8
254	0.0	0.2 0.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.428 min Scan# 1904


Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Tgt Ion:152	Resp:	130358
Ion Ratio	Lower	Upper
152	100	
115	57.7	28.8 86.5
150	163.7	0.0 348.8









#73
Isopropylbenzene
Concen: 20.315 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:105 Resp: 167050
Ion Ratio Lower Upper
105 100
120 27.0 13.3 39.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#74
N-ethyl acetate
Concen: 19.251 ug/l
RT: 12.148 min Scan# 1694
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43


Tgt Ion: 43 Resp: 49888
Ion Ratio Lower Upper
43 100
70 46.5 29.7 44.5#
55 26.1 18.3 27.5
61 26.0 17.5 26.3





#75
1,1,2,2-Tetrachloroethane
Concen: 19.843 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.006 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43




Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01



Tgt Ion: 83 Resp: 42309
Ion Ratio Lower Upper
83 100
131 10.0 5.4 16.2
85 64.8 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#76
1,2,3-Trichloropropane
Concen: 18.676 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 75 Resp: 28421
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 20.142 ug/l

RT: 12.611 min Scan# 1770

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01





Tgt Ion:	156	Resp:	4061
Ion Ratio	Lower	Upper	
156	100		
77	189.3	92.8	278.6
158	96.1	48.4	145.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#78

n-propylbenzene

Concen: 20.650 ug/l




RT: 12.672 min Scan# 1780


Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Tgt Ion:	91	Resp:	204206
Ion Ratio	Lower	Upper	
91	100		
120	22.8	11.5	34.4





#79

2-Chlorotoluene

Concen: 20.631 ug/l

RT: 12.758 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01

Tgt Ion: 91 Resp: 11621

Ion Ratio Lower Upper

91 100





126 35.1 17.3 51.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#80

1,3,5-Trimethylbenzene

Concen: 20.675 ug/l

RT: 12.812 min Scan# 1803

Delta R.T. 0.000 min

Lab File: VY015916.D




Acq: 11 Oct 2023 11:43


Tgt Ion:105 Resp: 140125

Ion Ratio Lower Upper

105 100

120 49.6 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 18.110 ug/l
RT: 12.380 min Scan# 1732
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 75 Resp: 12428
Ion Ratio Lower Upper
75 100
53 91.6 77.8 116.8
89 47.7 37.4 56.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#82
4-Chlorotoluene
Concen: 20.523 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 91 Resp: 120471
Ion Ratio Lower Upper
91 100
126 34.9 17.2 51.6









#83
tert-Butylbenzene
Concen: 20.954 ug/l
RT: 13.075 min Scan# 119
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:119 Resp: 132534
Ion Ratio Lower Upper
119 100
91 62.6 32.9 98.6
134 25.3 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#84
1,2,4-Trimethylbenzene
Concen: 20.556 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:105 Resp: 140296
Ion Ratio Lower Upper
105 100
120 45.5 22.2 66.6





#85

sec-Butylbenzene

Concen: 20.740 ug/l

RT: 13.258 min Scan# 1876

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01

Tgt Ion:105 Resp: 192389

Ion Ratio Lower Upper

105 100





134 20.6 10.2 30.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#86

p-Isopropyltoluene

Concen: 20.873 ug/l

RT: 13.373 min Scan# 1895

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43




Tgt Ion:119 Resp: 160172


Ion Ratio Lower Upper

119 100

134 26.4 13.3 39.8


91 22.5 12.4 37.4





#87
 1,3-Dichlorobenzene
 Concen: 20.561 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY015916.D
 Acq: 11 Oct 2023 11:43




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1011SBSD01




Tgt Ion:146 Resp: 83468
 Ion Ratio Lower Upper
 146 100
 111 38.7 19.4 58.2
 148 64.2 31.9 95.5

Manual Integrations
 APPROVED



Reviewed By :John Carlone 10/12/2023
 Supervised By :Mahesh Dadoda 10/12/2023




#88
 1,4-Dichlorobenzene
 Concen: 20.435 ug/l
 RT: 13.447 min Scan# 1907
 Delta R.T. 0.000 min
 Lab File: VY015916.D
 Acq: 11 Oct 2023 11:43



Tgt Ion:146 Resp: 83539
 Ion Ratio Lower Upper
 146 100
 111 40.0 18.9 56.7
 148 64.3 32.1 96.5









#89
n-Butylbenzene
Concen: 21.221 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion: 91 Resp: 15410
Ion Ratio Lower Upper
91 100
92 53.7 25.4 76.0
134 25.9 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#90
Hexachloroethane
Concen: 20.485 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:117 Resp: 31547
Ion Ratio Lower Upper
117 100
201 75.7 41.6 124.8









#91
1,2-Dichlorobenzene
Concen: 20.694 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:146 Resp: 7808
Ion Ratio Lower Upper
146 100
111 41.5 19.7 59.0
148 63.2 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.440 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion: 75 Resp: 6905
Ion Ratio Lower Upper
75 100
155 92.1 41.5 124.6
157 127.6 51.8 155.5





#93

1,2,4-Trichlorobenzene

Concen: 20.376 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. 0.000 min

Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Instrument :

MSVOA_Y

ClientSampleId :

VY1011SBSD01

Tgt Ion:180 Resp: 49275





Ion	Ratio	Lower	Upper
180	100		
182	95.4	47.0	141.2
145	30.8	14.5	43.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/12/2023

Supervised By :Mahesh Dadoda 10/12/2023



#94

Hexachlorobutadiene

Concen: 20.759 ug/l

RT: 15.117 min Scan# 2181




Delta R.T. 0.006 min


Lab File: VY015916.D

Acq: 11 Oct 2023 11:43

Tgt Ion:225 Resp: 26882

Ion	Ratio	Lower	Upper
225	100		
223	64.4	31.9	95.9
227	65.3	33.0	98.9









#95
Naphthalene
Concen: 19.365 ug/l
RT: 15.239 min Scan# 2201
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Instrument :
MSVOA_Y
ClientSampleId :
VY1011SBSD01

Tgt Ion:128 Resp: 10833
Ion Ratio Lower Upper
128 100
127 12.7 10.8 16.2
129 11.1 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/12/2023
Supervised By :Mahesh Dadoda 10/12/2023



#96
1,2,3-Trichlorobenzene
Concen: 20.415 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.000 min
Lab File: VY015916.D
Acq: 11 Oct 2023 11:43

Tgt Ion:180 Resp: 43917
Ion Ratio Lower Upper
180 100
182 95.0 47.3 141.8
145 32.6 15.4 46.2

