



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101123\
Data File : VY015912.D
Acq On : 11 Oct 2023 08:42
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100923S.M
Title : SW846 8260
Last Update : Tue Oct 10 07:06:18 2023



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1742

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	22261	PASS
75	95	30	60	51.4	63379	PASS
95	95	100	100	100.0	123205	PASS
96	95	5	9	6.7	8267	PASS
173	174	0.00	2	1.0	1022	PASS
174	95	50	100	85.7	105600	PASS
175	174	5	9	7.5	7869	PASS
176	174	95	101	96.4	101837	PASS
177	176	5	9	6.8	6955	PASS