

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101222\
 Data File : VY010945.D
 Acq On : 12 Oct 2022 16:26
 Operator : KP/MD
 Sample : N5105-01
 Misc : 6.45g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-7

Quant Time: Oct 13 05:35:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	4302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	8569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	8968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3177	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	7287	167.109	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	334.220%#
35) Dibromofluoromethane	7.716	113	3550	62.410	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	124.820%
50) Toluene-d8	10.179	98	12203	62.731	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	125.460%
62) 4-Bromofluorobenzene	12.483	95	4020	53.125	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.260%
Target Compounds						Qvalue
20) Methylene Chloride	4.710	84	3172	67.146	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101222\
Data File : VY010945.D
Acq On : 12 Oct 2022 16:26
Operator : KP/MD
Sample : N5105-01
Misc : 6.45g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-7

Quant Time: Oct 13 05:35:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

