

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101223\
 Data File : VY015945.D
 Acq On : 12 Oct 2023 19:54
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Oct 13 13:14:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	162258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	272213	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	252720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	120127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	99674	53.825	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.660%	
35) Dibromofluoromethane	7.728	113	87591	51.111	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.220%	
50) Toluene-d8	10.185	98	322200	50.204	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.400%	
62) 4-Bromofluorobenzene	12.489	95	112190	49.093	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 98.180%	
Target Compounds						
3) Chloromethane	2.119	50	2271	Below Cal	Qvalue	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101223\
Data File : VY015945.D
Acq On : 12 Oct 2023 19:54
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Oct 13 13:14:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

