

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101321\
 Data File : VY006380.D
 Acq On : 13 Oct 2021 12:05
 Operator : SY/MD
 Sample : VY1013SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1013SBL01

Quant Time: Oct 14 07:04:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	143149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	219826	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	199477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	97190	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	76049	49.637	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.280%
35) Dibromofluoromethane	7.722	113	55631	47.368	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.740%
50) Toluene-d8	10.185	98	227393	46.263	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.520%
62) 4-Bromofluorobenzene	12.483	95	72250	43.269	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101321\
Data File : VY006380.D
Acq On : 13 Oct 2021 12:05
Operator : SY/MD
Sample : VY1013SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1013SBL01

Quant Time: Oct 14 07:04:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

