

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101321\
 Data File : VY006380.D
 Acq On : 13 Oct 2021 12:05
 Operator : SY/MD
 Sample : VY1013SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1013SBL01

Quant Time: Oct 14 07:04:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	143149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	219826	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	199477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	97190	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	76049	49.637	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.280%
35) Dibromofluoromethane	7.722	113	55631	47.368	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.740%
50) Toluene-d8	10.185	98	227393	46.263	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.520%
62) 4-Bromofluorobenzene	12.483	95	72250	43.269	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.540%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101321\
Data File : VY006380.D
Acq On : 13 Oct 2021 12:05
Operator : SY/MD
Sample : VY1013SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1013SBL01

Quant Time: Oct 14 07:04:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 980
Delta R.T. -0.000 min
Lab File: VY006380.D
Acq: 13 Oct 2021 12:05




Instrument :
MSVOA_Y
ClientSampleId :
VY1013SBL01


Tgt Ion:168 Resp: 143149
Ion Ratio Lower Upper
168 100
99 54.1 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 49.637 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. -0.006 min
Lab File: VY006380.D
Acq: 13 Oct 2021 12:05

Tgt Ion: 65 Resp: 76049
Ion Ratio Lower Upper
65 100
67 49.5 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.006 min

Lab File: VY006380.D

Acq: 13 Oct 2021 12:05





Tgt Ion:114 Resp: 219826

Ion Ratio Lower Upper

114 100

63 21.9 0.0 43.2

88 15.0 0.0 31.8



#35

Dibromofluoromethane

Concen: 47.368 ug/l

RT: 7.722 min Scan# 968

Delta R.T. 0.000 min

Lab File: VY006380.D

Acq: 13 Oct 2021 12:05




Tgt Ion:113 Resp: 55631


Ion Ratio Lower Upper

113 100

111 102.6 82.6 124.0

192 23.3 17.6 26.4









#50
Toluene-d8
Concen: 46.263 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY006380.D
Acq: 13 Oct 2021 12:05




Instrument :
MSVOA_Y
ClientSampleId :
VY1013SBL01


Tgt Ion: 98 Resp: 227393
Ion Ratio Lower Upper
98 100
100 62.5 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 43.269 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY006380.D
Acq: 13 Oct 2021 12:05

Tgt Ion: 95 Resp: 72250
Ion Ratio Lower Upper
95 100
174 95.0 0.0 177.0
176 91.2 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.495 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VY006380.D
Acq: 13 Oct 2021 12:05

Instrument :
MSVOA_Y
ClientSampleId :
VY1013SBL01

Tgt Ion:	117	Resp:	199477
Ion	Ratio	Lower	Upper
117	100		
82	53.7	43.0	64.4
119	31.3	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY006380.D
Acq: 13 Oct 2021 12:05

Tgt Ion:	152	Resp:	97190
Ion	Ratio	Lower	Upper
152	100		
115	55.3	28.2	84.7
150	156.6	0.0	347.6

