

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
 Data File : VY015954.D
 Acq On : 13 Oct 2023 13:44
 Operator : SY/MD
 Sample : 04858-04
 Misc : 5.71g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2

Quant Time: Oct 14 01:49:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	139313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	239877	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	230263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	111003	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	92002	57.865	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.720%
35) Dibromofluoromethane	7.716	113	81488	53.960	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.920%
50) Toluene-d8	10.179	98	285815	50.538	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.080%
62) 4-Bromofluorobenzene	12.483	95	100497	49.904	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.800%
Target Compounds						
20) Methylene Chloride	4.710	84	18455	6.884	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
Data File : VY015954.D
Acq On : 13 Oct 2023 13:44
Operator : SY/MD
Sample : 04858-04
Misc : 5.71g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2

Quant Time: Oct 14 01:49:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

