

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
 Data File : VY015957.D
 Acq On : 13 Oct 2023 15:07
 Operator : SY/MD
 Sample : 04844-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-603-COMP-06

Quant Time: Oct 13 16:08:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	13229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	22385	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	23211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	11846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	10715	70.970	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.940%
35) Dibromofluoromethane	7.728	113	8367	59.372	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.740%
50) Toluene-d8	10.185	98	26156	49.561	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.120%
62) 4-Bromofluorobenzene	12.489	95	10459	55.655	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	111.320%
Target Compounds						
16) Acetone	3.966	43	1613	18.392	ug/l #	47
20) Methylene Chloride	4.722	84	2631	12.606	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
Data File : VY015957.D
Acq On : 13 Oct 2023 15:07
Operator : SY/MD
Sample : 04844-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-603-COMP-06

Quant Time: Oct 13 16:08:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

