

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
 Data File : VY015959.D
 Acq On : 13 Oct 2023 15:52
 Operator : SY/MD
 Sample : 04844-01RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-603-COMP-05RE

Quant Time: Oct 13 17:46:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	47597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	76879	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	71359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	32759	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	29911	55.063	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.120%
35) Dibromofluoromethane	7.722	113	26352	54.447	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.900%
50) Toluene-d8	10.185	98	90049	49.682	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.360%
62) 4-Bromofluorobenzene	12.489	95	30362	47.043	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.080%
Target Compounds						
20) Methylene Chloride	4.710	84	5394	5.235	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
Data File : VY015959.D
Acq On : 13 Oct 2023 15:52
Operator : SY/MD
Sample : 04844-01RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-603-COMP-05RE

Quant Time: Oct 13 17:46:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

