

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
 Data File : VY015960.D
 Acq On : 13 Oct 2023 16:16
 Operator : SY/MD
 Sample : 04844-13
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-603-COMP-08

Quant Time: Oct 13 17:54:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	46751	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	69530	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	60668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	25626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	23525	44.091	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.180%
35) Dibromofluoromethane	7.728	113	22623	51.683	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.360%
50) Toluene-d8	10.185	98	80304	48.988	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.980%
62) 4-Bromofluorobenzene	12.489	95	24266	41.572	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.140%
Target Compounds						
20) Methylene Chloride	4.723	84	6376	7.221	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
Data File : VY015960.D
Acq On : 13 Oct 2023 16:16
Operator : SY/MD
Sample : 04844-13
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-603-COMP-08

Quant Time: Oct 13 17:54:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

