

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
 Data File : VY015964.D
 Acq On : 13 Oct 2023 17:49
 Operator : SY/MD
 Sample : 04878-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11995

Quant Time: Oct 14 01:50:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	9493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	15206	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	14222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	6097	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	6756	62.358	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.720%
35) Dibromofluoromethane	7.728	113	5730	59.856	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.720%
50) Toluene-d8	10.191	98	17836	49.752	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.500%
62) 4-Bromofluorobenzene	12.489	95	5721	44.816	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
Data File : VY015964.D
Acq On : 13 Oct 2023 17:49
Operator : SY/MD
Sample : 04878-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11995

Quant Time: Oct 14 01:50:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

