

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
 Data File : VY015969.D
 Acq On : 13 Oct 2023 19:47
 Operator : SY/MD
 Sample : 04888-01
 Misc : 4.72g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-304-COMP-01

Quant Time: Oct 14 01:52:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	130221	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.703	114	217936	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	201770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	94151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	81807	55.045	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.100%
35) Dibromofluoromethane	7.734	113	70966	51.724	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.440%
50) Toluene-d8	10.191	98	258881	50.384	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.760%
62) 4-Bromofluorobenzene	12.489	95	87338	47.736	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.480%
Target Compounds						
20) Methylene Chloride	4.735	84	25545	12.373	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101323\
Data File : VY015969.D
Acq On : 13 Oct 2023 19:47
Operator : SY/MD
Sample : 04888-01
Misc : 4.72g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-304-COMP-01

Quant Time: Oct 14 01:52:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

