

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101623\
 Data File : VY015981.D
 Acq On : 16 Oct 2023 13:55
 Operator : SY/MD
 Sample : 04798-01
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FRIABLE-ASBESTOS-SAMPLE

Quant Time: Oct 17 05:42:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	137030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	234061	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	225666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	109269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	84468	54.011	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.020%	
35) Dibromofluoromethane	7.722	113	67250	45.638	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 91.280%	
50) Toluene-d8	10.184	98	284650	51.583	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 103.160%	
62) 4-Bromofluorobenzene	12.489	95	104110	52.983	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 105.960%	
Target Compounds						Qvalue
16) Acetone	3.984	43	14046	15.461	ug/l	93
20) Methylene Chloride	4.716	84	27241	12.599	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101623\
Data File : VY015981.D
Acq On : 16 Oct 2023 13:55
Operator : SY/MD
Sample : 04798-01
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FRIABLE-ASBESTOS-SAMPLE

Quant Time: Oct 17 05:42:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

