

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101722\
 Data File : VY010974.D
 Acq On : 17 Oct 2022 16:32
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY101722

Quant Time: Oct 18 01:45:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.332	0.320	3.6	104	0.00
3 P	Chloromethane	0.476	0.463	2.7	107	0.00
4 C	Vinyl Chloride	0.539	0.540	-0.2#	107	0.00
5 T	Bromomethane	0.373	0.356	4.6	109	0.00
6 T	Chloroethane	0.360	0.361	-0.3	108	0.00
7 T	Trichlorofluoromethane	0.789	0.792	-0.4	107	0.00
8 T	Diethyl Ether	0.295	0.285	3.4	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.493	0.4	106	0.00
10 T	Methyl Iodide	0.606	0.629	-3.8	104	0.00
11 T	Tert butyl alcohol	0.062	0.048	22.6	101	-0.01
12 CM	1,1-Dichloroethene	0.474	0.469	1.1#	105	0.00
13 T	Acrolein	0.074	0.066	10.8	93	0.00
14 T	Allyl chloride	0.752	0.752	0.0	104	0.00
15 T	Acrylonitrile	0.154	0.146	5.2	99	0.00
16 T	Acetone	0.134	0.149	-11.2	118	0.00
17 T	Carbon Disulfide	1.391	1.305	6.2	103	0.00
18 T	Methyl Acetate	0.428	0.353	17.5	100	0.00
19 T	Methyl tert-butyl Ether	1.325	1.317	0.6	102	0.00
20 T	Methylene Chloride	0.761	0.575	24.4	104	0.00
21 T	trans-1,2-Dichloroethene	0.535	0.532	0.6	105	0.00
22 T	Diisopropyl ether	1.611	1.655	-2.7	104	0.00
23 T	Vinyl Acetate	0.963	1.007	-4.6	102	0.00
24 P	1,1-Dichloroethane	0.973	0.971	0.2	104	0.00
25 T	2-Butanone	0.203	0.198	2.5	103	0.00
26 T	2,2-Dichloropropane	0.864	0.864	0.0	106	0.00
27 T	cis-1,2-Dichloroethene	0.617	0.618	-0.2	104	0.00
28 T	Bromochloromethane	0.303	0.361	-19.1	100	0.00
29 T	Tetrahydrofuran	0.125	0.119	4.8	98	0.00
30 C	Chloroform	0.980	0.985	-0.5#	104	0.00
31 T	Cyclohexane	0.881	0.831	5.7	104	0.00
32 T	1,1,1-Trichloroethane	0.862	0.861	0.1	104	0.00
33 S	1,2-Dichloroethane-d4	0.461	0.420	8.9	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.304	0.269	11.5	100	0.00
36 T	1,1-Dichloropropene	0.465	0.462	0.6	104	0.00
37 T	Ethyl Acetate	0.258	0.253	1.9	101	0.00
38 T	Carbon Tetrachloride	0.480	0.482	-0.4	104	0.00
39 T	Methylcyclohexane	0.557	0.563	-1.1	105	0.00
40 TM	Benzene	1.366	1.349	1.2	103	0.00
41 T	Methacrylonitrile	0.140	0.129	7.9	105	0.00
42 TM	1,2-Dichloroethane	0.378	0.376	0.5	103	0.00
43 T	Isopropyl Acetate	0.460	0.454	1.3	101	0.00
44 TM	Trichloroethene	0.370	0.366	1.1	104	0.00
45 C	1,2-Dichloropropane	0.347	0.346	0.3#	103	0.00
46 T	Dibromomethane	0.193	0.194	-0.5	104	0.00
47 T	Bromodichloromethane	0.466	0.470	-0.9	103	0.00
48 T	Methyl methacrylate	0.206	0.213	-3.4	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101722\
 Data File : VY010974.D
 Acq On : 17 Oct 2022 16:32
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY101722

Quant Time: Oct 18 01:45:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	98	0.00
50 S	Toluene-d8	1.044	0.996	4.6	99	0.00
51 T	4-Methyl-2-Pentanone	0.257	0.251	2.3	99	0.00
52 CM	Toluene	0.848	0.859	-1.3#	104	0.00
53 T	t-1,3-Dichloropropene	0.484	0.488	-0.8	103	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.558	-0.7	102	0.00
55 T	1,1,2-Trichloroethane	0.279	0.278	0.4	102	0.00
56 T	Ethyl methacrylate	0.370	0.379	-2.4	101	0.00
57 T	1,3-Dichloropropane	0.479	0.473	1.3	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.164	0.176	-7.3	94	0.00
59 T	2-Hexanone	0.176	0.180	-2.3	102	0.00
60 T	Dibromochloromethane	0.332	0.329	0.9	102	0.00
61 T	1,2-Dibromoethane	0.267	0.260	2.6	101	0.00
62 S	4-Bromofluorobenzene	0.418	0.350	16.3	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.389	0.386	0.8	102	0.00
65 PM	Chlorobenzene	1.027	1.036	-0.9	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.383	-0.8	103	0.00
67 C	Ethyl Benzene	1.816	1.872	-3.1#	103	0.00
68 T	m/p-Xylenes	0.699	0.719	-2.9	104	0.00
69 T	o-Xylene	0.663	0.686	-3.5	103	0.00
70 T	Styrene	1.124	1.164	-3.6	102	0.00
71 P	Bromoform	0.234	0.235	-0.4	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.633	3.769	-3.7	104	0.00
74 T	N-amyl acetate	0.922	0.941	-2.1	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.762	0.740	2.9	100	0.00
76 T	1,2,3-Trichloropropane	0.568	0.574	-1.1	104	0.00
77 T	Bromobenzene	0.849	0.861	-1.4	103	0.00
78 T	n-propylbenzene	4.420	4.598	-4.0	103	0.00
79 T	2-Chlorotoluene	2.489	2.551	-2.5	103	0.00
80 T	1,3,5-Trimethylbenzene	3.031	3.157	-4.2	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.264	0.267	-1.1	100	0.00
82 T	4-Chlorotoluene	2.592	2.648	-2.2	103	0.00
83 T	tert-Butylbenzene	2.621	2.726	-4.0	102	0.00
84 T	1,2,4-Trimethylbenzene	2.989	3.095	-3.5	103	0.00
85 T	sec-Butylbenzene	3.938	4.127	-4.8	103	0.00
86 T	p-Isopropyltoluene	3.267	3.452	-5.7	104	0.00
87 T	1,3-Dichlorobenzene	1.695	1.717	-1.3	103	0.00
88 T	1,4-Dichlorobenzene	1.685	1.684	0.1	102	0.00
89 T	n-Butylbenzene	3.053	3.263	-6.9	105	0.00
90 T	Hexachloroethane	0.645	0.662	-2.6	104	0.00
91 T	1,2-Dichlorobenzene	1.516	1.513	0.2	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.120	4.8	99	0.00
93 T	1,2,4-Trichlorobenzene	0.906	0.944	-4.2	104	0.00
94 T	Hexachlorobutadiene	0.533	0.548	-2.8	104	0.00
95 T	Naphthalene	1.845	1.917	-3.9	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101722\
Data File : VY010974.D
Acq On : 17 Oct 2022 16:32
Operator : KP/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY101722

Quant Time: Oct 18 01:45:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.791	0.815	-3.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6