



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101722\
 Data File : VY010965.D
 Acq On : 17 Oct 2022 09:00
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Title : SW846 8260
 Last Update : Tue Oct 18 01:42:42 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	27773	PASS
75	95	30	60	50.2	79293	PASS
95	95	100	100	100.0	158059	PASS
96	95	5	9	6.6	10472	PASS
173	174	0.00	2	0.9	1177	PASS
174	95	50	100	82.5	130411	PASS
175	174	5	9	7.4	9706	PASS
176	174	95	101	95.4	124411	PASS
177	176	5	9	6.6	8256	PASS