

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101723\
 Data File : VY016001.D
 Acq On : 17 Oct 2023 14:00
 Operator : SY/MD
 Sample : 04892-01RE
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-02-10132023-ARE

Quant Time: Oct 17 15:56:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	74057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	121482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	114097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	56629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	44329	52.448	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.900%	
35) Dibromofluoromethane	7.722	113	41866	54.741	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.480%	
50) Toluene-d8	10.185	98	142386	49.714	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.420%	
62) 4-Bromofluorobenzene	12.489	95	49152	48.195	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 96.400%	
Target Compounds						Qvalue
80) 1,3,5-Trimethylbenzene	12.818	105	6825	2.049	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101723\
Data File : VY016001.D
Acq On : 17 Oct 2023 14:00
Operator : SY/MD
Sample : 04892-01RE
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-02-10132023-ARE

Quant Time: Oct 17 15:56:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

