

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101723\
 Data File : VY016007.D
 Acq On : 17 Oct 2023 16:21
 Operator : SY/MD
 Sample : 04905-01
 Misc : 4.67g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LSS-B4-COMP-101323

Quant Time: Oct 18 06:09:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	120961	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	207279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	190640	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	91529	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	76318	55.283	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.560%
35) Dibromofluoromethane	7.722	113	69513	53.269	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.540%
50) Toluene-d8	10.179	98	243337	49.794	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.580%
62) 4-Bromofluorobenzene	12.483	95	80379	46.191	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.380%
Target Compounds						
16) Acetone	4.003	43	10361	12.920	ug/l	99
17) Carbon Disulfide	4.186	76	10553	2.935	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101723\
Data File : VY016007.D
Acq On : 17 Oct 2023 16:21
Operator : SY/MD
Sample : 04905-01
Misc : 4.67g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LSS-B4-COMP-101323

Quant Time: Oct 18 06:09:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration

