

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101723\
 Data File : VY016004.D
 Acq On : 17 Oct 2023 15:11
 Operator : SY/MD
 Sample : 04908-01
 Misc : 4.64g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1

Quant Time: Oct 18 06:08:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 13 13:13:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	128228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	214194	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	199517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	97974	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	75783	51.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.560%
35) Dibromofluoromethane	7.722	113	69897	51.834	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.660%
50) Toluene-d8	10.185	98	251404	49.784	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.560%
62) 4-Bromofluorobenzene	12.483	95	85740	47.681	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.360%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101723\
Data File : VY016004.D
Acq On : 17 Oct 2023 15:11
Operator : SY/MD
Sample : 04908-01
Misc : 4.64g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1

Quant Time: Oct 18 06:08:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101223S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 13 13:13:02 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 980
Delta R.T. -0.000 min
Lab File: VY016004.D
Acq: 17 Oct 2023 15:11




Instrument :
MSVOA_Y
ClientSampleId :
TP-1


Tgt Ion:168 Resp: 128228
Ion Ratio Lower Upper
168 100
99 52.4 43.4 65.0



#33
1,2-Dichloroethane-d4
Concen: 51.784 ug/l
RT: 8.148 min Scan# 1038
Delta R.T. -0.000 min
Lab File: VY016004.D
Acq: 17 Oct 2023 15:11

Tgt Ion: 65 Resp: 75783
Ion Ratio Lower Upper
65 100
67 55.6 0.0 101.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. -0.006 min
Lab File: VY016004.D
Acq: 17 Oct 2023 15:11




Instrument :
MSVOA_Y
ClientSampleId :
TP-1


Tgt Ion:114 Resp: 214194
Ion Ratio Lower Upper
114 100
63 19.0 0.0 42.4
88 14.7 0.0 30.6



#35
Dibromofluoromethane
Concen: 51.834 ug/l
RT: 7.722 min Scan# 968
Delta R.T. -0.000 min
Lab File: VY016004.D
Acq: 17 Oct 2023 15:11

Tgt Ion:113 Resp: 69897
Ion Ratio Lower Upper
113 100
111 102.5 81.3 121.9
192 19.4 16.9 25.3









#50
Toluene-d8
Concen: 49.784 ug/l
RT: 10.185 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY016004.D
Acq: 17 Oct 2023 15:11




Instrument :
MSVOA_Y
ClientSampleId :
TP-1


Tgt Ion: 98 Resp: 251404
Ion Ratio Lower Upper
98 100
100 65.1 50.1 75.1



#62
4-Bromofluorobenzene
Concen: 47.681 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY016004.D
Acq: 17 Oct 2023 15:11

Tgt Ion: 95 Resp: 85740
Ion Ratio Lower Upper
95 100
174 85.5 0.0 178.2
176 82.2 0.0 173.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.495 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY016004.D

Acq: 17 Oct 2023 15:11

Instrument :

MSVOA_Y

ClientSampleId :

TP-1





Tgt Ion:117 Resp: 199517

Ion Ratio Lower Upper

117 100

82 53.0 43.4 65.0

119 30.7 26.3 39.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.428 min Scan# 1904

Delta R.T. -0.000 min

Lab File: VY016004.D

Acq: 17 Oct 2023 15:11

Tgt Ion:152 Resp: 97974

Ion Ratio Lower Upper

152 100

115 57.6 28.8 86.5

150 157.6 0.0 348.8

