

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101724\
 Data File : VY019928.D
 Acq On : 17 Oct 2024 13:41
 Operator : SY/MD
 Sample : P4425-05
 Misc : 5.53g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: Oct 18 01:40:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	239321	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	491804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	442546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	156184	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	165349	56.112	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.220%
35) Dibromofluoromethane	7.640	113	163231	50.297	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.600%
50) Toluene-d8	10.109	98	601562	50.194	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.380%
62) 4-Bromofluorobenzene	12.408	95	185264	42.782	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	85.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101724\
Data File : VY019928.D
Acq On : 17 Oct 2024 13:41
Operator : SY/MD
Sample : P4425-05
Misc : 5.53g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: Oct 18 01:40:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

