

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
 Data File : VY010987.D
 Acq On : 18 Oct 2022 11:51
 Operator : KP/MD
 Sample : N5105-01
 Misc : 5.81g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-7

Quant Time: Oct 18 15:23:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	60177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	103185	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	93043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	41506	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	65436	118.048	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 236.100%#	
35) Dibromofluoromethane	7.716	113	54840	97.386	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 194.780%#	
50) Toluene-d8	10.179	98	166098	77.093	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 154.180%#	
62) 4-Bromofluorobenzene	12.483	95	53267	71.091	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 142.180%	
Target Compounds						Qvalue
16) Acetone	3.942	43	3630	22.550	ug/l #	80
18) Methyl Acetate	4.473	43	1159	Below Cal	#	74
20) Methylene Chloride	4.716	84	20553	26.304	ug/l #	84
93) 1,2,4-Trichlorobenzene	15.001	180	783	1.041	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
Data File : VY010987.D
Acq On : 18 Oct 2022 11:51
Operator : KP/MD
Sample : N5105-01
Misc : 5.81g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-7

Quant Time: Oct 18 15:23:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

