

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
 Data File : VY010992.D
 Acq On : 18 Oct 2022 13:44
 Operator : KP/MD
 Sample : N5149-11
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-VPB-RW5-GW-938-940

Quant Time: Oct 18 17:03:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	237430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	389428	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	327737	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	110216	50.394	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.780%	
35) Dibromofluoromethane	7.716	113	115739	53.466	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.940%	
50) Toluene-d8	10.179	98	407594	50.127	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.260%	
62) 4-Bromofluorobenzene	12.477	95	126648	43.575	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 87.160%	
Target Compounds						
16) Acetone	3.942	43	3674	5.785	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
Data File : VY010992.D
Acq On : 18 Oct 2022 13:44
Operator : KP/MD
Sample : N5149-11
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-VPB-RW5-GW-938-940

Quant Time: Oct 18 17:03:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

