

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
 Data File : VY010994.D
 Acq On : 18 Oct 2022 14:33
 Operator : KP/MD
 Sample : N5133-01
 Misc : 4.97g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HAS-SOIL-1013

Quant Time: Oct 18 17:24:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	326414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	561521	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	490335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	212861	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	153346	51.001	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.000%
35) Dibromofluoromethane	7.716	113	159023	50.841	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.680%
50) Toluene-d8	10.179	98	551587	47.045	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.100%
62) 4-Bromofluorobenzene	12.483	95	183278	43.745	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
Data File : VY010994.D
Acq On : 18 Oct 2022 14:33
Operator : KP/MD
Sample : N5133-01
Misc : 4.97g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HAS-SOIL-1013

Quant Time: Oct 18 17:24:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

