

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
 Data File : VY010983.D
 Acq On : 18 Oct 2022 09:44
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 18 15:21:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.332	0.316	4.8	95	0.00
3 P	Chloromethane	0.476	0.458	3.8	98	0.00
4 C	Vinyl Chloride	0.539	0.540	-0.2#	98	0.00
5 T	Bromomethane	0.373	0.361	3.2	102	0.00
6 T	Chloroethane	0.360	0.360	0.0	99	0.00
7 T	Trichlorofluoromethane	0.789	0.776	1.6	97	0.00
8 T	Diethyl Ether	0.295	0.286	3.1	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.484	2.2	96	0.00
10 T	Methyl Iodide	0.606	0.626	-3.3	95	0.00
11 T	Tert butyl alcohol	0.062	0.044	29.0#	86	0.00
12 CM	1,1-Dichloroethene	0.474	0.461	2.7#	95	0.00
13 T	Acrolein	0.074	0.066	10.8	85	0.00
14 T	Allyl chloride	0.752	0.758	-0.8	97	0.00
15 T	Acrylonitrile	0.154	0.149	3.2	93	0.00
16 T	Acetone	0.134	0.143	-6.7	104	0.00
17 T	Carbon Disulfide	1.391	1.312	5.7	95	0.00
18 T	Methyl Acetate	0.428	0.352	17.8	92	0.00
19 T	Methyl tert-butyl Ether	1.325	1.306	1.4	93	0.00
20 T	Methylene Chloride	0.761	0.557	26.8#	93	0.00
21 T	trans-1,2-Dichloroethene	0.535	0.531	0.7	96	0.00
22 T	Diisopropyl ether	1.611	1.657	-2.9	96	0.00
23 T	Vinyl Acetate	0.963	1.008	-4.7	94	0.00
24 P	1,1-Dichloroethane	0.973	0.980	-0.7	97	0.00
25 T	2-Butanone	0.203	0.194	4.4	93	0.00
26 T	2,2-Dichloropropane	0.864	0.860	0.5	97	0.00
27 T	cis-1,2-Dichloroethene	0.617	0.625	-1.3	97	0.00
28 T	Bromochloromethane	0.303	0.366	-20.8	93	0.00
29 T	Tetrahydrofuran	0.125	0.119	4.8	91	0.00
30 C	Chloroform	0.980	0.996	-1.6#	97	0.00
31 T	Cyclohexane	0.881	0.821	6.8	95	0.00
32 T	1,1,1-Trichloroethane	0.862	0.863	-0.1	96	0.00
33 S	1,2-Dichloroethane-d4	0.461	0.429	6.9	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.304	0.279	8.2	95	0.00
36 T	1,1-Dichloropropene	0.465	0.468	-0.6	97	0.00
37 T	Ethyl Acetate	0.258	0.246	4.7	90	0.00
38 T	Carbon Tetrachloride	0.480	0.485	-1.0	97	0.00
39 T	Methylcyclohexane	0.557	0.556	0.2	95	0.00
40 TM	Benzene	1.366	1.367	-0.1	96	0.00
41 T	Methacrylonitrile	0.140	0.141	-0.7	106	0.01
42 TM	1,2-Dichloroethane	0.378	0.375	0.8	94	0.00
43 T	Isopropyl Acetate	0.460	0.450	2.2	92	0.00
44 TM	Trichloroethene	0.370	0.369	0.3	97	0.00
45 C	1,2-Dichloropropane	0.347	0.355	-2.3#	97	0.00
46 T	Dibromomethane	0.193	0.194	-0.5	96	0.00
47 T	Bromodichloromethane	0.466	0.473	-1.5	95	0.00
48 T	Methyl methacrylate	0.206	0.205	0.5	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
 Data File : VY010983.D
 Acq On : 18 Oct 2022 09:44
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 18 15:21:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	88	0.00
50 S	Toluene-d8	1.044	1.024	1.9	94	0.00
51 T	4-Methyl-2-Pentanone	0.257	0.249	3.1	90	0.00
52 CM	Toluene	0.848	0.863	-1.8#	96	0.00
53 T	t-1,3-Dichloropropene	0.484	0.487	-0.6	95	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.563	-1.6	95	0.00
55 T	1,1,2-Trichloroethane	0.279	0.280	-0.4	95	0.00
56 T	Ethyl methacrylate	0.370	0.374	-1.1	91	0.00
57 T	1,3-Dichloropropane	0.479	0.479	0.0	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.164	0.180	-9.8	89	0.00
59 T	2-Hexanone	0.176	0.177	-0.6	92	0.00
60 T	Dibromochloromethane	0.332	0.334	-0.6	95	0.00
61 T	1,2-Dibromoethane	0.267	0.260	2.6	93	0.00
62 S	4-Bromofluorobenzene	0.418	0.359	14.1	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.389	0.377	3.1	93	0.00
65 PM	Chlorobenzene	1.027	1.031	-0.4	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.385	-1.3	97	0.00
67 C	Ethyl Benzene	1.816	1.848	-1.8#	96	0.00
68 T	m/p-Xylenes	0.699	0.714	-2.1	97	0.00
69 T	o-Xylene	0.663	0.684	-3.2	96	0.00
70 T	Styrene	1.124	1.158	-3.0	95	0.00
71 P	Bromoform	0.234	0.232	0.9	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.633	3.806	-4.8	96	0.00
74 T	N-amyl acetate	0.922	0.917	0.5	90	0.00
75 P	1,1,2,2-Tetrachloroethane	0.762	0.754	1.0	94	0.00
76 T	1,2,3-Trichloropropane	0.568	0.613	-7.9	102	0.00
77 T	Bromobenzene	0.849	0.861	-1.4	95	0.00
78 T	n-propylbenzene	4.420	4.632	-4.8	96	0.00
79 T	2-Chlorotoluene	2.489	2.570	-3.3	96	0.00
80 T	1,3,5-Trimethylbenzene	3.031	3.174	-4.7	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.264	0.268	-1.5	92	0.00
82 T	4-Chlorotoluene	2.592	2.662	-2.7	95	0.00
83 T	tert-Butylbenzene	2.621	2.731	-4.2	94	0.00
84 T	1,2,4-Trimethylbenzene	2.989	3.131	-4.8	95	0.00
85 T	sec-Butylbenzene	3.938	4.130	-4.9	95	0.00
86 T	p-Isopropyltoluene	3.267	3.429	-5.0	95	0.00
87 T	1,3-Dichlorobenzene	1.695	1.702	-0.4	94	0.00
88 T	1,4-Dichlorobenzene	1.685	1.703	-1.1	95	0.00
89 T	n-Butylbenzene	3.053	3.247	-6.4	96	0.00
90 T	Hexachloroethane	0.645	0.669	-3.7	96	0.00
91 T	1,2-Dichlorobenzene	1.516	1.518	-0.1	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.118	6.3	90	0.00
93 T	1,2,4-Trichlorobenzene	0.906	0.914	-0.9	92	0.00
94 T	Hexachlorobutadiene	0.533	0.542	-1.7	95	0.00
95 T	Naphthalene	1.845	1.866	-1.1	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101822\
Data File : VY010983.D
Acq On : 18 Oct 2022 09:44
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 18 15:21:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.791	0.787	0.5	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6