

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101823\
 Data File : VY016029.D
 Acq On : 18 Oct 2023 18:52
 Operator : SY/MD
 Sample : 04938-02
 Misc : 5.31g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A-1(2-6)

Quant Time: Oct 19 03:27:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 19 03:15:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	139241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	239864	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	219070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	90074	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	94308	59.592	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.180%
35) Dibromofluoromethane	7.722	113	80161	50.773	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.540%
50) Toluene-d8	10.185	98	277596	48.412	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.820%
62) 4-Bromofluorobenzene	12.483	95	88593	43.301	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101823\
Data File : VY016029.D
Acq On : 18 Oct 2023 18:52
Operator : SY/MD
Sample : 04938-02
Misc : 5.31g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A-1(2-6)

Quant Time: Oct 19 03:27:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 19 03:15:44 2023
Response via : Initial Calibration

