

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101823\
 Data File : VY016030.D
 Acq On : 18 Oct 2023 19:15
 Operator : SY/MD
 Sample : 04888-01
 Misc : 5.65g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-304-COMP-21

Quant Time: Oct 19 03:28:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 19 03:15:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	142074	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	243471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	225519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	106011	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	95692	59.260	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.520%
35) Dibromofluoromethane	7.728	113	82166	51.272	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.540%
50) Toluene-d8	10.185	98	286191	49.172	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.340%
62) 4-Bromofluorobenzene	12.489	95	97074	46.743	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.480%
Target Compounds						Qvalue
20) Methylene Chloride	4.722	84	13895	7.814	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101823\
Data File : VY016030.D
Acq On : 18 Oct 2023 19:15
Operator : SY/MD
Sample : 04888-01
Misc : 5.65g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-304-COMP-21

Quant Time: Oct 19 03:28:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 19 03:15:44 2023
Response via : Initial Calibration

