

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101921\
 Data File : VY006417.D
 Acq On : 19 Oct 2021 10:37
 Operator : SY/MD
 Sample : VY1019SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1019SBL01

Quant Time: Oct 20 03:14:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	124057	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	193069	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	176740	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87162	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	70642	53.20	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.40%
35) Dibromofluoromethane	7.722	113	52349	50.75	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.50%
50) Toluene-d8	10.185	98	208857	48.38	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.76%
62) 4-Bromofluorobenzene	12.483	95	68521	46.72	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.44%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101921\
Data File : VY006417.D
Acq On : 19 Oct 2021 10:37
Operator : SY/MD
Sample : VY1019SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1019SBL01

Quant Time: Oct 20 03:14:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

