

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101923\
 Data File : VY016038.D
 Acq On : 19 Oct 2023 15:36
 Operator : SY/MD
 Sample : 04919-01
 Misc : 5.77g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LSS-B2-COMP-101723

Quant Time: Oct 20 05:37:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 19 03:15:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	151032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	256072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	235565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	111158	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	97596	56.855	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.700%
35) Dibromofluoromethane	7.716	113	83623	49.613	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.220%
50) Toluene-d8	10.179	98	299998	49.008	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.020%
62) 4-Bromofluorobenzene	12.483	95	100484	46.004	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.000%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	20532	10.862	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101923\
Data File : VY016038.D
Acq On : 19 Oct 2023 15:36
Operator : SY/MD
Sample : 04919-01
Misc : 5.77g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LSS-B2-COMP-101723

Quant Time: Oct 20 05:37:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 19 03:15:44 2023
Response via : Initial Calibration

